AMENDMENT NO. 3 OCTOBER 2018
TO
IS 7080 (PART 3) : 1992 MTP SUCTION APPARATUS — SPECIFICATION

PART 3 ELECTRICALLY OPERATED

(*Page* 4, *clause* **10.4**) — Substitute the following for the existing clause:

'**10.4** For Type A apparatus, the unit shall be capable of giving a vacuum of 66 percent of the atmospheric pressure at that place in 10 s and 79 percent of the atmospheric pressure at that place within 60 s with a minimum free air displacement rate of 25 l/min. The unit shall be mounted in such a way as to produce minimum vibrations.

For Type B apparatus, the unit shall be capable of giving a vacuum of 66 percent of the atmospheric pressure at that place in 10 s and 92 percent of the atmospheric pressure at that place within 50 s with a minimum free displacement rate of 25 to 30 l/min. The vacuum unit shall be explosion proof and shall not produce sparks during operation. It shall be mounted in a such a way as to produce minimum vibrations.'

(MHD 03)

Publication Unit, BIS, New Delhi, India